



## Certification of Conformity

EU-MEDICAL DEVICES DIRECTIVE-93/42/EEC

Registration No.: ENC161014GZ71

Applicant Address : No. 5-2, Wudayuan Road, East Lake High Tech Development Zone, Wuhan, China

Product Designation : Cranial Electrotherapy Stimulation (CES) Device

Model Number : AT-9, AT-11

The submitted products have been tested by us with the listed standard and found in compliance with the Medical Devices Directive 93/42/EEC and following European Standards:

Directive	Applied Standards	Test Report No.
EMC Directive	EN 60601-1-2:2015	ENC161014GZ71E1
	EN 61000-3-2:2014	
	EN 61000-3-3:2013	
Safety Directive	EN 60601-1:2006+A11:2011+A1:2013 +A12:2014	ENC161014GZ71L1

This certificate is based on an evaluation of a sample of the above mentioned product. Technical report and documentation are at the licence applicant's disposal. This certificate does not imply assessment of the series-production of the product. The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.



Ray Zhou / General Manager  
Issue Date: Dec. 7, 2016

**East Notice Certification Service Co., Ltd.**

1/F, Haohui Commercial Building, Zhuji Street, Dongpu Town, Tianhe District, Guangzhou City  
Tel:+86-020-2331 4234 Fax:+86-020-8256 8534 E-mail:enc@enc-lab.com Web: www.enc-lab.com