DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

MANUFACTURER:	CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA	
MEDICAL DEVICE:	Electrocardiograph ECG600G	
CLASSIFICATION - ANNEX IX:	Class II a, Rule 10	
CONFORMITY ASSESSMENT ROUTE:	Annex II excluding chapter 4	
MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; NCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE. STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.		
NOTIFIED BODY:	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 M NCHEN, GERMANY	
IDENTIFICATION NUMBER:	CE 0123	
(EC) CERTIFICATE(S):	G1 050972 0050 Rev.03	
EUROPEAN REPRESENTATIVE:	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany	
START OF CE-MARKING:	2012-04-20 (Date or Lot or serial number)	

PLACE, DATE OF DECLARATION:	Qinhuangdao, 2019-11-07		
SIGNATURE:	- B R M Pre	esident	
		TF-CE101216-09	Ver: I
	J	Page 1 of 2	

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

No.	Reference	Title of Standard
1	IEC 60601-1: 1988	Medical electrical equipment; Part 1: General requirements
	+A1:1991+A2:1995	for safety
2	IEC 60601-1-6:2006	Medical electrical equipment Part 1-6: General
		requirements for safety - Collateral Standard: Usability
3	IEC60601-2-25: 1993 +	Medical electrical equipment – Part 2-25: Particular
	A1:1999	requirements for the safety of electrocardiographs
EN60601 1 4:100	EN60601-1-4:1996	Medical electrical equipment; Part 1: General requirements
4	+A1:1999	for safety –4 Collateral standard: Programmable electrical
	TAI.1999	medical systems
		Medical electrical equipment Part 1: General requirements
5	IEC 60601-1-2: 2007	for safety -2 Collateral standard: Electromagnetic
		compatibility - Requirements and tests
6	IEC62304:2006	Medical device software Software life cycle processes

TF-CE101216-09	Ver: I
Page 2 of 2	