

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**



MANUFACTURER:

CONTEC MEDICAL SYSTEMS CO., LTD
No.112 Qinhuang West Street, Economic & Technical
Development Zone, Qinhuangdao, Hebei Province,
PEOPLE'S REPUBLIC OF CHINA

MEDICAL DEVICE:

Ambulatory Blood Pressure Monitor, ABPM50

CLASSIFICATION - ANNEX IX:

Class II a, Rule 10

CONFORMITY ASSESSMENT ROUTE: Annex II excluding chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED
MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF
COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;
INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC
ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH
DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER:

CE 0123

(EC) CERTIFICATE(S):

G1 050972 0050 Rev.02

EC REP

EUROPEAN REPRESENTATIVE:

Shanghai International Holding Corp. GmbH(Europe)
Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING:

2011-01-14 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:

QINHUANGDAO, 2019-07-23

SIGNATURE:



President

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	IEC 60601-1: 2012	Medical electrical equipment- Part1: General requirements for basic safety and essential performance
2	EN 60601-1-2: 2007	Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN 62366:2008 (IEC 62366:2007)	Medical devices - Application of usability engineering to medical devices
4	EN 60601-1-6:2010	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
5	IEC80601-2-30:2009+A1:2013 (Edition1.1)	Medical electrical equipment –Part 2-30: Particular requirements for the basic safety and essential performance of automatic non-invasive sphygmomanometers
6	EN 62304:2006	Medical device software –Software life cycle processes