

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES


MANUFACTURER:	CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
MEDICAL DEVICE:	Patient Monitor , CMS8000
CLASSIFICATION - ANNEX IX:	Class II b, Rule 10
CONFORMITY ASSESSMENT ROUTE:	Annex II excluding chapter 4
GMDN CODE:	33586

We, (CONTEC MEDICAL SYSTEMS CO., LTD) herewith declare that the stated medical devices meet the transposition into national law, the provisions of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices;
Including, at 21 March 2010, the amendments by Council Directive 2007/47/EEC
All supporting documentation is retained at the premises of the manufacture.

Standards applied: see attached list of (harmonised - EN) standards for which documented evidence of compliance can be provided.

NOTIFIED BODY:	TÜV SÜD Product service GmbH Ridlerstr 65, D-80339 München, Germany
IDENTIFICATION NUMBER:	 0123
(EC) CERTIFICATE(S):	<u>G1 050972 0050 Rev.04</u>
EUROPEAN REPRESENTATIVE:	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING: 2010-3-20 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:	QINHUANGDAO, 2020-06-18
SIGNATURE:	 _____ President

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

No.	Serial Number	Title and Description
1	IEC 62366-1:2015	Medical devices - Application of usability engineering to medical devices
2	IEC 62304:2015	Medical device software - Software life-cycle processes
3	IEC 60601-1:2012	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
4	IEC60601-1-2:2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
5	IEC 60601-1-6:2013	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability
6	IEC60601-1-8:2012	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
7	IEC 60601-2-27:2011	Medical electrical equipment - Part 2-27: Particular requirements for the safety, including essential performance, of electrocardiographic monitoring equipment
8	IEC 80601-2-30:2018	Medical electrical equipment -Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers
9	IEC 60601-2-34:2011	Medical electrical equipment -Part 2-34: Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment
10	IEC 80601-2-49:2018	Medical electrical equipment - Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitoring equipment
11	ISO 80601-2-55:2018	Medical electrical equipment Part 2: Particular requirements for the basic safety and essential performance of respiratory gas monitors
12	ISO 80601-2-56:2017 +A1: 2018	Medical electrical equipment - Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement
13	ISO 80601-2-61:2017	Medical electrical equipment -Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment